



CASE REPORT

Life threatening ventilatory obstruction due to a foreign body in a new sealed endotracheal tubePadara Basavaraj¹, Jamkar MA², Vishwanath Shettar³¹ Assistant Professor Department of Anaesthesiology, BJ Medical College and Sassoon Hospital Pune, Maharashtra, India.² Associate Professor Department of Anaesthesiology, BJ Medical College and Sassoon Hospital Pune, Maharashtra, India.³ Assistant Professor Department of Anaesthesiology, BJ Medical College and Sassoon Hospital Pune, Maharashtra, India.**Received 23 October 2014; Accepted 28 October 2014****INTRODUCTION:**

In the current days of high tech equipment and well defined safety regulations, technical failures are less likely to occur. Still many manufacturing and packaging errors are encountered in anaesthetic practice which go unnoticed during the routine inspection prior to their use. A 32 year old healthy adult male with insignificant medical and surgical history was posted for clavicular plating under general endotracheal anaesthesia (GETA). Inside the operation theatre premedication was given after attaching monitors, patient was preoxygenated for 3 minutes with 100% oxygen. After induction with thiopentone sodium and confirming ventilation, succinyl choline was administered for muscle relaxation. Laryngoscopy was done, a new portex cuffed endotracheal tube of 8.5 size was inserted under vision and connected to the breathing circuit. On ventilation after cuff inflation there was no chest expansion and air entry on auscultation. High resistance to manual ventilation along with above mentioned findings made us to think of some obstruction to the ventilation, and we checked the breathing system for any kinks or obstructions from common gas outlet till endotracheal tube and we could not find anything of that sort. Then we made a possible diagnosis of acute severe bronchospasm and initiated our measures towards relieving the bronchospasm in a stepwise manner according to the algorithm for managing perioperative bronchospasm (1). In spite of all measures taken there was no improvement in chest expansion and air entry, and oxygen saturation (Spo₂) then started decreasing gradually to 50%. Then it was decided to do repeat laryngoscopy to confirm the position of endotracheal tube, and it was found to be

properly positioned. when we decided to reintubate the patient and were taking preparations, we saw some white colored plastic piece of foreign material inside the ETT connector extending into lumen of the tube (figure 1). Immediately connector was removed and piece of foreign material removed from it and connected to ETT, and ventilated with 100% oxygen. There was clear air entry and adequate chest expansion with oxygen saturation gradually increasing to 100% within 30 sec. procedure went well and patient extubated at end without any complications.

Common reasons for failure to ventilate after successful endotracheal intubation are anaesthetic gas delivery malfunction, breathing system obstruction, acute severe bronchospasm, poor pulmonary compliance, tension pneumothorax etc. Some uncommon observations have also been reported till date and include, endobronchial mass lesion (2), cuff herniation leading to intraluminal tracheal obstruction(3,4), elliptical defects in the wall causing air leak(5), intraluminal plastic film and meniscus(6), plastic film with a small central perforation covering distal end of ETT(7), ETT connector defect(8) etc...

Routine pre-operative check up of ETT should be a standard of practice, ETT should be checked for patency of the lumen along with pilot ballon assembly. In spite of routine pre use check up some defects go unnoticed and are evident only after initiation of mechanical ventilation. Our observation highlights the importance of double checking of anaesthetic equipment before use and using standard monitoring of mechanical ventilation. Thus the role of vigilant anaesthetist is invaluable for patient safety.



Figure 1:



Figure 1:

REFERENCES:

1. Looseley Alex .Management of bronchospasm during general anaesthesia.Updates in anaesthesia, www.anaesthesiologist.com.2011;17-21
2. Barst S.Yossefy Y.Lebowitz P.An unusual cause of airway obstruction.Anaesth Analg 1994;78:195
3. Henderson MA. Airway obstruction with a cuffed single use plastic endotracheal tube .Anaesth Intensive care 1993;21:370-2
4. Famewo CE. A not so apparent cause of intraluminal tracheal tube obdtruction Anaesthesiology 1983;58: 593
5. Lower BM, Karim Z, Henderson RS. Large air leak from an endotracheal tube due to a manufacturing defect. Anaesth analg 1997;85:944-5
6. Ngan Kee WD. An unusual problem with an endotracheal tube.Anaesth Intensive Care 1993;21:247-8
7. Hajimohammadi F, Taheri A,Eghtensadi Araghi P.Obstruction of endotracheal tube a manufacturing error. Middle east j Anaesthesiol 2009;20:303-5
8. Shamsberry C,Kannauja A, Gautam S. Ventilation failur due to endotracheal T -connector defect .Indian J Anaesth.2010 July-aug 54(4):357-8